

MethaMeasure User Guide

(MAN003 For use in prison environments)

Helpdesk: 0800 027 0671



1	Description	3
2	Intended Use	3
3	Warnings	4
4	Symbols	4
5	Logging In	5
6	Setup the Pump(s).....	5
7	Calibrate the Pump.....	6
8	Tips for Pump Calibration.....	7
9	Add a new Patient	8
10	Change Patient Details	8
11	How to take Patients Photograph	8
12	How to Capture a Patient Fingerprint	8
13	Adding a prescription	9
14	How to change a prescription (mid-prescription).....	10
15	How to delete a prescription.....	10
16	Dispensing a Patient.....	10
17	Managing Stock	11
18	Entering Stock	11
19	Making Corrections	11
20	Important Information	12
20.1	View / Print previous Registers.....	12
20.2	If the pump runs out of Methadone while dispensing a dose (i.e the tubes have air bubbles in them).....	12
20.3	Return doses to stock (if patient has been dispensed by not collected).....	12
20.4	Correcting errors on the register	13
20.5	Close Down Procedure.....	13
20.6	Cleaning the Pump.....	13
20.7	Tablet Dispensing.....	14
20.8	Shutdown Procedure	15
20.9	Changing the tube/ service volume	16
21.	Keyboard Shortcut table	18

1 Description

The MethaMeasure System is composed of the MethaMeasure software (installed on a computer) and a liquid pump, intended to accurately dispense doses of a liquid drug (Methadone) from a stock container (2.5 or 5l).

The system is shown in Figure 1 – a computer on the right, printer on the left, and a double pump cabinet. The double pump cabinet is an option for regular methadone on one side, and sugar free methadone on the other side. Selection of regular or sugar free is configured for each patient, all other functionality is identical. The pump is a rotary peristaltic pump, driven by a stepper motor for accurate control of rotation, and hence metering of the fluid in the tubing fitted into the pump. Silicone tubing of a known Internal Diameter (ID) of 4.8mm (for 1mg/ml) can be replaced by the user when the tubing is worn.

The side compartments of the cabinet have the stock container, and these compartments are locked to prevent access to the Methadone. The dispensing container (usually a cup but could be a bottle for advance dispensing) is placed below the outlet tube in the central compartment.

Methadone is a controlled drug, requiring accurate control and records of dispensing. The accurate measurement of dispensing is the medical device function, and the subject of this technical documentation. The software on the computer enforces a daily calibration procedure, where a known volume is dispensed, and the user must measure and report on this volume. Once calibrated, the software can dispense accurate volumes. Management of the drug stock, patients and prescriptions are not medical device functions, and these functions of the software are out of scope.



Figure 1. The MethaMeasure System with a Small Double UK Standard Pump (MMUKP02).

2 Intended Use








The MethaMeasure system device is intended for use by a pharmacist to accurately dispense a liquid drug that must be controlled (requirement by pharmaceutical regulations). Only the accurate dispensing is within the scope of this intended use.



3 Warnings

1. The user must report any serious incident that has occurred in relation to the medical device to the manufacturer and the relevant Competent Authority having jurisdiction in their locale.
2. The user must install tubing correctly, refer to section 20.9
3. Only use tubing supplied by MethaMeasure Ltd in the pump.
4. The user must clean the device: refer to section 20.6.

4 Symbols

	Manufacturer
	Authorised representative in the European Community/ Union
	Serial Number
	Medical Device
	Unique Device Identifier
	Consult instructions for use
	Software version

5 Logging In

1. Double click on the MethaMeasure Icon on the Desktop
2. Log in both members of staff, first the Nurse/Supervisor role, and then the second checker/Operator role

OR add a new member of staff by clicking '**Add a user**'

6 Setup the Pump(s)




Once the Main MethaMeasure window appears remove any bottles of cleaning fluid from the machine.

1. Turn the Pump ON and place a measure or cup under the dispensing nozzle. Then click the "options" button on the pump status box
2. Click 'Prime Pump' until all excess water is removed from the tubes. **Once** should be enough.
3. Insert the Methadone bottle into the pump and make sure the tube is correctly positioned. Then prime the pump **once** again into same container. The tube will be filled with Methadone

***** Throw away this water/methadone mix *****

4. Place the glass measure under the dispensing nozzle and click 'Prime Pump' one last time to check there are no air bubbles in the tube.
***** Pour this methadone back into stock. *****

You have now primed the pump



7 Calibrate the Pump

Click the "Calibration" button

1. The machine will ask to dispense 100ml, this should be dispensed into the glass measure used for MethaMeasure

2. The machine will then ask the user to "Check Volume", Enter the volume dispensed into the glass measure into the box and click **next**. Return this methadone to stock.




3. The machine will repeat the dose if the amount is NOT 100ml.

Once the 100ml is OK the machine will move onto the 50ml and 10ml doses.

4. When you have finished calibration close the calibration menu to continue to the main window.

The pump you have just calibrated will turn GREEN.

The pump is now ready for dispensing.





8 Tips for Pump Calibration

Calibrating the pump accurately to a glass measure is an important step, if the pump is not accurately calibrated the dose given to a patient is likely to be incorrect. Whilst an inaccuracy of 0.5-1% is unlikely to cause a patient any harm it will cause a large discrepancy in the running balance over a high number of patients. Some tips for calibration include

- Use a good quality glass measure, we recommend a cylindrical measure from Jaytec with an accuracy of +/- 0.1-0.5ml. These can be obtained at a cost from MethaMeasure. Please call for pricing. However, anything with the quality assurance stamp and glass should be suitable. We do NOT recommend plastic.
- Below is a table of recommended tube sizes and glassware sizes based on Methadone Strength. Please contact us before using anything other than 1mg/ml in our pumps.

Methadone Strength	Recommended Tube Size	Glassware Size/Accuracy
10mg/ml	14 – 1.6mm ID	10ml/0.1ml
5mg/ml	14 – 1.6mm ID	10ml/0.1ml
1mg/ml	25 – 4.8mm ID	100ml/0.5ml

- Use the glass measure in the priming process so it is “wet” before you start the calibration. It must be carefully drained after priming however it will help keep the measure consistent through the calibration process
- Use just one measure for the calibration and the same measure every day if you can.
- Carefully drain as much Methadone after calibration back into the stock bottle as possible – if you leave, for example, 1-2ml in the bottom of the measure and dispense 50ml on top – it will look like 51-52ml.
- Hold the measure at an angle whilst the calibration dose is being dispensed
- Do not allow the nozzle to become fully immersed in the Methadone whilst the dose is being dispensed – this will lead to some methadone being left on the dispensing tube which will make the measured dose inaccurate and also cause drips.
- Read the bottom of the meniscus on a flat and level surface
- Make sure the pump is fully primed, with zero air bubbles in the tubes, before starting calibration. A common mistake we see is people not priming the tube, meaning the calibration has to fill the pump first which results in 85ml or so being dispensed during the first calibration.



9 Add a new Patient

1. Click on "Add Patient" Icon (plus symbol) at the top left of the screen
2. Fill in the form with the patient details shown on the prescription.
3. Choose if the patient is Supervised, or not
4. Click '**Save Patient**'

Add patient

Name

Date of Birth (optional) (dd/MM/yyyy)

Address

Phone number (optional)

Postcode (optional)

Supervised Active


Save patient

10 Change Patient Details

1. Select the patient you want to change.
2. Click '**Edit details**'
3. Make any changes required and click '**Save Patient**'


11 How to take Patients Photograph

1. Select the patient you want to add a Picture for
2. Click '**Edit Image**'
3. Make sure the patient can be seen clearly in the image
4. Click '**Capture**' to take the image
5. Click '**Save**'



12 How to Capture a Patient Fingerprint

1. Select the patient you want to add a Fingerprint
2. Select '**Edit fingerprint**' on the menu to the right
3. Ask the patient to place their finger on the sensor until the capture is successful.
4. Ask patient to remove their finger and then to place it back on the sensor.
5. Click '**Save**'



13 Adding a prescription

1. Select the patient you wish to add a script for
2. Click the **'Add Prescription'** button on the right of the patient profile
3. Enter the information as it appears on the Prescription paying careful attention to the start date. Carefully check the daily dose to be given and click **'Next'**
4. To save this prescription click **'Confirm'**
5. The script will now be visible in the patient profile

New prescription for Steve Smith

Prescriber: Jones Clinic (optional): Titration dosing
 Drug: Physopstone SF Individual dose: 20 mg (20 ml)
 Drug strength: 1 mg/ml
 Total dose (mg): 280 Total days: 14
 Start date: 08/03/2021 End date: 20/03/2021

New prescription for John Smith

Mon	Tue	Wed	Thu	Fri	Sat	Sun
00:00	00:00	00:00	00:00	00:00	00:00	00:00
00:00	00:00	00:00	00:00	00:00	00:00	00:00
00:00	00:00	00:00	00:00	00:00	00:00	00:00


700 Methadone SF
 Keep prescribed drug
50 Individual dose mg
700 Total in calendar mg

John Smith
 A13213AS
 Supervised

Date	Ref	Drug	Dose	Amount	Type	Status
5 Sep 2017	01	Methadone SF 1 mg/ml	50 mg	50 ml	Daily	available
6 Sep 2017	02	Methadone SF 1 mg/ml	50 mg	50 ml	Daily	future
7 Sep 2017	03	Methadone SF 1 mg/ml	50 mg	50 ml	Daily	future
8 Sep 2017	04	Methadone SF 1 mg/ml	50 mg	50 ml	Daily	future
9 Sep 2017	05	Methadone SF 1 mg/ml	50 mg	50 ml	Daily	future
10 Sep 2017	06	Methadone SF 1 mg/ml	50 mg	50 ml	Daily	future
11 Sep 2017	07	Methadone SF 1 mg/ml	50 mg	50 ml	Daily	future
12 Sep 2017	08	Methadone SF 1 mg/ml	50 mg	50 ml	Daily	future
13 Sep 2017	09	Methadone SF 1 mg/ml	50 mg	50 ml	Daily	future
14 Sep 2017	010	Methadone SF 1 mg/ml	50 mg	50 ml	Daily	future
15 Sep 2017	011	Methadone SF 1 mg/ml	50 mg	50 ml	Daily	future
16 Sep 2017	012	Methadone SF 1 mg/ml	50 mg	50 ml	Daily	future
17 Sep 2017	013	Methadone SF 1 mg/ml	50 mg	50 ml	Daily	future
18 Sep 2017	014	Methadone SF 1 mg/ml	50 mg	50 ml	Daily	future

14 How to change a prescription (mid-prescription)

1. Select the patient that you want to change the prescription for.
2. Click '**Prescriptions & doses**' on the menu to the right.
3. Select the Prescription you wish to edit
4. Click '**Cancel**' on the right of the screen
5. Click '**Add Prescription**' at the bottom of the screen.
6. Enter new prescription as shown (see adding a prescription).



15 How to delete a prescription

1. Select the patient that you want to change the prescription for.
 2. Click '**Prescriptions & doses**' on the menu to the right.
 3. Select the Prescription you wish to edit
 4. Click '**Delete**' on the right of the screen
- * You can now go back and re enter the prescription as above



16 Dispensing a Patient

To dispense a dose for a patient, click on the patients name and choose "Begin dispensing".

The details can be checked against the prescription, click "Pump now" to dispense the dose using the pump, print a label and make the entry in the CD register.

"Mark dispensed" can be used to label the dose and make an entry on the CD register for the dose if it the pump is not available and the dose has been hand measured or the drug you are dispensing is not allocated to a pump.







17 Managing Stock

All methadone stock is managed by clicking the **'Stock'** button on the main screen.

1. Select 'Store Stock' and select which drug you wish to manage

2. This will now show you all the stock information about that drug.



18 Entering Stock

1. Click **'Book stock in'**

2. Enter the amount of stock you have received and any other information you wish to record and click **'Save'**

19 Making Corrections

1. Select **'Make Correction'**



2. Type the amount that you wish to correct
* Enter '40' for an Overage or '-40' for reducing stock

3. Select the reason for the correction from the dropdown menu

4. Click **'Save'**

Record Spillages

*See making Corrections





20 Important Information

If you have problems with the machine please look through this list. If you still have problems you can contact our help desk on:

0800 027 0671 Option 1

The helpline is open from **08:45 to approximately 18:00 Monday to Saturday**.
Out of hours service is available on a Sunday.

Request Help button

If your request is *non urgent* you can send a quick message to the help desk by clicking on the support button at the top right of the MethaMeasure screen

20.1 View / Print previous Registers

1. Click '**Registers**' on the main menu
2. Select which drug you wish to view
3. Chose the date you wish to view at the top of the screen
4. Click '**Print Day**'

20.2 If the pump runs out of Methadone while dispensing a dose (i.e the tubes have air bubbles in them)

1. Firstly return the methadone in the cup to the stock bottle
2. Refill the Methadone bottle with a bottle from the CD cabinet
3. Click '**Prescriptions & Doses**'
4. Click the '**View Doses**' Button at the bottom of the screen
5. Select the dose you wish to undo **click 'Return Selected Dose'**

Select the reason for returning the dose from the drop down menu

6. Click '**Return Selected doses(s)**' button
7. Press **Esc** key until you are back at the patient screen
8. Click the '**Prime**' button to refill the tube with Methadone.

***Return any methadone to stock**

9. Select the patient and re dispense

20.3 Return doses to stock (if patient has been dispensed by not collected)

1. Click patient name.
2. **Click 'Prescriptions & Doses'**
3. Click the '**View Doses**' Button at the bottom of the screen
4. Select the dose you wish to undo **click 'Return Selected Dose'**
5. Select the reason for returning the dose from the drop down menu
6. Click '**Return Selected doses(s)**' button



20.4 Correcting errors on the register

If errors are identified on the register they can only be corrected via '**Manual Correction**' under the '**Stock**' menu.

20.5 Close Down Procedure

After Dispensing has finished the methadone pump needs to be cleaned out and the register(s) checked.

20.6 Cleaning the Pump

1. Remove any bottles of Methadone from the machine.
2. Place a measure or cup under the nozzle, Click Options on the pump status screen and then click '**Prime Pump**' until all the Methadone is removed from the tubes. – Pour this Methadone back into stock and return the Methadone to CD Cabinet
3. Insert cleaning fluid bottle into the pump and make sure the tube is correctly positioned.

Place a measure or cup under the nozzle.

4. Click 'Prime Pump' until you can see water in the tube. Click Prime Pump until the tube is clean. Once a week we recommend a diluted Milton solution is used to flush through the system as a disinfectant. Please do not leave the Milton Solution in the pump – use the prime feature to wash the Milton solution through the pump and then refill with fresh water.

****Leave this water in the tube overnight**


5. The pump can be turned off if required.

You have now cleaned out the pump

20.7 Tablet Dispensing

MethaMeasure can be used to record the dispensing of tablets (And also to help with ID checks). Tablets are handling in a similar way to Methadone. Stock is monitored through the stock menu and registers are accessed through the register menu. If you need help with adding Tablets as drugs to MethaMeasure please contact the help desk.

Add a prescription in the usual way, choosing the tablet name as the drug and entering the **total dose** in mg to be given. In this case 140mg over 14 days, click **Next** and choose the take home days, Click **Confirm**



New prescription for Bob Anyname

Prescriber: Jones Clinic (optional):

Drug: Buprenorphine

Total dose (mg): 140 Total days: 14

Start date: 08/03/2021 End date: 21/03/2021

You can choose the tablet quantities for each dose after the next step.

Titration dosing

Individual dose: 10 mg

→ Next

↩ Cancel

Choose the quantities of each tablet strength to be given to the patient. This mechanism allows a patient to be given multiple of each tablet strength in order to get to their required dose. In this case, 0x0.4mg, 1x2mg and 1x8mg, however, it could be 0x 0.4mg and 5x2mg and 0x 8mg's if there were no 8mg tablets in stock. Click **Next** when finished.



New prescription for Bob Anyname

Enter the quantities of each strength for each tablet dose below.

	0.4 mg	2 mg	8 mg
10/10 mg	0	1	1

→ Next

↩ Cancel

Dispense the tablet patient as normal, when **begin dispensing** is clicked the dispensing screen will be displayed. A list of each strength of tablet will be displayed. It is very important to dispense the tablets **exactly** as prescribed on the system otherwise stock checks will not be correct. Click **Mark Dispensed** to record the patient as dispensed.





20.8 Shutdown Procedure

1. Close MethaMeasure by clicking the 'X' in the top right
2. A box will pop up asking you to review today's register.
3. Click '**View Registers**'
4. Check that the register entries are correct and that all collections and non collections are registered correctly. And click '**Continue Review**' button

**If you have a double pump click on the other drug register to be checked.*

6. Once the registers are checked click the '**Complete**' button.
7. Check the box next to the final confirmation that the registers have been checked and are accurate.
8. Click the '**Confirm**' button
9. If you choose to skip this process, an entry is made on the register to say the entries have not been checked.




20.9 Changing the tube/ service volume

The silicone tube used in the MethaMeasure peristaltic pump is a wearing component and must be replaced at appropriate intervals. Depending on the tube in use on your pump and the volume of Methadone dispensed by the system this may require changing every 3 months or it may last considerably longer. As the tube comes to the end of its service life the MethaMeasure software will alert you to this by changing the colour of the pump controller and displaying a message in the pump status. If you do not have a spare tube on site, please call us on 0800 027 0671 and we will send you an appropriate tube. This is included in your rental of MethaMeasure.

As the tube nears the end of its life, the status will change to "Ready, Change hose"



A warning will be displayed depending on the remaining hose/tube life



If the warning is ignored, the system will eventually stop dispensing once it has reached a predetermined volume.




To change the tube, first use the prime button to empty the Methadone from the system. Then change the tube, please see separate document, "MethaMeasure UK Tubing Change MAN002" available from our website www.methameasure.co.uk under the support button if required.

Once the tube has been changed click the pump options menu and prime the pump. Then click the pump servicing menu and click "Reset service volume", the system will ask the user to confirm they have changed the hose. The user will then be taken



through the calibration process (see section 6 page 5 for priming and calibration instructions). Once calibration is complete the system can be used again.






21. Keyboard Shortcut table

MethaMeasure has several keyboard shortcuts to allow the software to be used without the mouse for many common tasks. Hovering over a button will prompt MethaMeasure to display a tooltip and also the keyboard shortcut if one is available. A summary is listed below.

Shortcut	Action
General Shortcuts	
Esc Key	Will move back a level in menus, or if on the main screen select the search box or clear the search box
Enter Key	Will move forwards a level as a positive menu action with a direction to dispense to the patient. Eg. Enter from the patient profile will move to the dispensing screen and Enter again will dispense the patient via the pump (if available)
Search Box	Searching can be by first name or surname in any order, once there is a single result on the main screen Enter will open the patient profile.
From the Main Screen	
Ctrl+F9	Opens the Stock Menu
Ctrl+F10	Opens the Register Menu
Ctrl+F11	Opens the Reports Menu
Ctrl+F12	Opens the Tools Menu
Ctrl+N	Add New patient
Ctrl+U	User menu/logout
Ctrl+P	Start fingerprint recognition
Ctrl+F	Select Search box
From within a Patient Profile	
Ctrl+N	View Notes
Ctrl+E	Edit Details
Ctrl+F	Add Fingerprint
Ctrl+I	Add Photograph
Ctrl+A	Add Prescription
Ctrl+P	Open Prescriptions and Doses
Ctrl+M	View Patient Messages (if enabled)
Ctrl+L	Print Bag Label



DOCUMENT DETAILS

 METHAMEASURE LIMITED	MethaMeasure User Guide - UK Community Pharmacy		
	No: MAN001	Rev: 01	Page: 16/16

REVISION HISTORY

Rev	CR #	Approval / Effective Date	Description of Change	Originator
0	CR-2023-08	7 December 2023	Release documentation	C Cave
01	CR-2024-03	14 June 2024	Added intended use, symbols, and warnings. Reformatted	C Cave